













Respironics SimplyGo	Product weight	Continuous flow settings	Pulse mode bolus size	Mobile Cart	Carrying Case	Extra Battery																				
	10lbs	<table border="1"> <thead> <tr> <th>Continuous Flow Settings</th> </tr> </thead> <tbody> <tr><td>0.5 LPM</td></tr> <tr><td>1.0 LPM</td></tr> <tr><td>1.5 LPM</td></tr> <tr><td>2.0 LPM</td></tr> </tbody> </table>	Continuous Flow Settings	0.5 LPM	1.0 LPM	1.5 LPM	2.0 LPM	<table border="1"> <thead> <tr> <th>Pulse Mode Settings</th> </tr> </thead> <tbody> <tr><td>1.0 = 12ml</td></tr> <tr><td>1.5 = 18ml</td></tr> <tr><td>2.0 = 24ml</td></tr> <tr><td>2.5 = 30ml</td></tr> <tr><td>3.0 = 36ml</td></tr> <tr><td>3.5 = 42ml</td></tr> <tr><td>4.0 = 48ml</td></tr> <tr><td>4.5 = 54ml</td></tr> <tr><td>5.0 = 60ml</td></tr> <tr><td>5.5 = 66ml</td></tr> <tr><td>6.0 = 72ml</td></tr> </tbody> </table>	Pulse Mode Settings	1.0 = 12ml	1.5 = 18ml	2.0 = 24ml	2.5 = 30ml	3.0 = 36ml	3.5 = 42ml	4.0 = 48ml	4.5 = 54ml	5.0 = 60ml	5.5 = 66ml	6.0 = 72ml	Yes 	Yes 	Yes 			
		Continuous Flow Settings																								
		0.5 LPM																								
1.0 LPM																										
1.5 LPM																										
2.0 LPM																										
Pulse Mode Settings																										
1.0 = 12ml																										
1.5 = 18ml																										
2.0 = 24ml																										
2.5 = 30ml																										
3.0 = 36ml																										
3.5 = 42ml																										
4.0 = 48ml																										
4.5 = 54ml																										
5.0 = 60ml																										
5.5 = 66ml																										
6.0 = 72ml																										
Battery Duration				AC Power Supply	DC Power Supply																					
		<table border="1"> <thead> <tr> <th>Flow Setting in Pulse on demand and constant flow</th> <th>Approx battery times with 1 battery installed ( Hrs: Minutes )</th> </tr> </thead> <tbody> <tr><td>Pulse 1 lpm</td><td>3.7</td></tr> <tr><td>Pulse 2 lpm</td><td>3.5</td></tr> <tr><td>Pulse 3 lpm</td><td>3.1</td></tr> <tr><td>Pulse 4 lpm</td><td>2.7</td></tr> <tr><td>Pulse 5 lpm</td><td>2.3</td></tr> <tr><td>Pulse 6 lpm</td><td>1.9</td></tr> <tr><td>Continuous 0.5 lpm</td><td>3.1</td></tr> <tr><td>Continuous 1.0 lpm</td><td>2.3</td></tr> <tr><td>Continuous 2.0 lpm</td><td>0.7</td></tr> </tbody> </table>	Flow Setting in Pulse on demand and constant flow	Approx battery times with 1 battery installed ( Hrs: Minutes )	Pulse 1 lpm	3.7	Pulse 2 lpm	3.5	Pulse 3 lpm	3.1	Pulse 4 lpm	2.7	Pulse 5 lpm	2.3	Pulse 6 lpm	1.9	Continuous 0.5 lpm	3.1	Continuous 1.0 lpm	2.3	Continuous 2.0 lpm	0.7		Yes 	Yes 	
Flow Setting in Pulse on demand and constant flow	Approx battery times with 1 battery installed ( Hrs: Minutes )																									
Pulse 1 lpm	3.7																									
Pulse 2 lpm	3.5																									
Pulse 3 lpm	3.1																									
Pulse 4 lpm	2.7																									
Pulse 5 lpm	2.3																									
Pulse 6 lpm	1.9																									
Continuous 0.5 lpm	3.1																									
Continuous 1.0 lpm	2.3																									
Continuous 2.0 lpm	0.7																									

Zen-0	Product weight	Continuous flow settings	Pulse mode bolus size	Mobile Cart	Carrying Case	Extra Battery																																																																		
	10.25lbs	<table border="1"> <thead> <tr> <th>Setting</th> <th>Flow rate</th> </tr> </thead> <tbody> <tr><td>0.5</td><td>0.5</td></tr> <tr><td>1.0</td><td>1.0</td></tr> <tr><td>1.5</td><td>1.5</td></tr> <tr><td>2.0</td><td>2.0</td></tr> </tbody> </table> <p><small>All values +/- 0.2 litres over all operating conditions</small></p>	Setting	Flow rate	0.5	0.5	1.0	1.0	1.5	1.5	2.0	2.0	<table border="1"> <thead> <tr> <th>Breath per minute</th> <th colspan="6">Setting</th> </tr> <tr> <th></th> <th>1</th> <th>2</th> <th>3</th> <th>4</th> <th>5</th> <th>6</th> </tr> </thead> <tbody> <tr><td>15</td><td>11</td><td>22</td><td>33</td><td>44</td><td>55</td><td>66</td></tr> <tr><td>20</td><td>11</td><td>22</td><td>33</td><td>44</td><td>55</td><td>66</td></tr> <tr><td>25</td><td>11</td><td>22</td><td>33</td><td>44</td><td>55</td><td>66</td></tr> <tr><td>30</td><td>11</td><td>22</td><td>33</td><td>44</td><td>55</td><td>66</td></tr> <tr><td>35</td><td>11</td><td>22</td><td>33</td><td>44</td><td>55</td><td>57</td></tr> <tr><td>40</td><td>11</td><td>22</td><td>33</td><td>44</td><td>50</td><td>50</td></tr> </tbody> </table> <p><small>All values +/- 15% over all operating conditions</small></p>	Breath per minute	Setting							1	2	3	4	5	6	15	11	22	33	44	55	66	20	11	22	33	44	55	66	25	11	22	33	44	55	66	30	11	22	33	44	55	66	35	11	22	33	44	55	57	40	11	22	33	44	50	50	Yes 	Yes 	Yes 
		Setting	Flow rate																																																																					
		0.5	0.5																																																																					
1.0	1.0																																																																							
1.5	1.5																																																																							
2.0	2.0																																																																							
Breath per minute	Setting																																																																							
	1	2	3	4	5	6																																																																		
15	11	22	33	44	55	66																																																																		
20	11	22	33	44	55	66																																																																		
25	11	22	33	44	55	66																																																																		
30	11	22	33	44	55	66																																																																		
35	11	22	33	44	55	57																																																																		
40	11	22	33	44	50	50																																																																		
Battery Duration				AC Power Supply	DC Power Supply																																																																			
		<table border="1"> <thead> <tr> <th>Flow Setting in Pulse on demand and constant flow</th> <th>Approx battery times with 1 battery installed ( Hrs: Minutes )</th> </tr> </thead> <tbody> <tr><td>Pulse 1 lpm</td><td>4</td></tr> <tr><td>Pulse 2 lpm</td><td>4</td></tr> <tr><td>Pulse 3 lpm</td><td>3</td></tr> <tr><td>Pulse 4 lpm</td><td>2.15</td></tr> <tr><td>Pulse 5 lpm</td><td>2</td></tr> <tr><td>Pulse 6 lpm</td><td>1.45</td></tr> <tr><td>Continuous 0.5 lpm</td><td>3</td></tr> <tr><td>Continuous 1.0 lpm</td><td>1.45</td></tr> <tr><td>Continuous 1.5 lpm</td><td>1.15</td></tr> <tr><td>Continuous 2.0 lpm</td><td>0.45</td></tr> </tbody> </table>	Flow Setting in Pulse on demand and constant flow	Approx battery times with 1 battery installed ( Hrs: Minutes )	Pulse 1 lpm	4	Pulse 2 lpm	4	Pulse 3 lpm	3	Pulse 4 lpm	2.15	Pulse 5 lpm	2	Pulse 6 lpm	1.45	Continuous 0.5 lpm	3	Continuous 1.0 lpm	1.45	Continuous 1.5 lpm	1.15	Continuous 2.0 lpm	0.45		Yes 	Yes 																																													
Flow Setting in Pulse on demand and constant flow	Approx battery times with 1 battery installed ( Hrs: Minutes )																																																																							
Pulse 1 lpm	4																																																																							
Pulse 2 lpm	4																																																																							
Pulse 3 lpm	3																																																																							
Pulse 4 lpm	2.15																																																																							
Pulse 5 lpm	2																																																																							
Pulse 6 lpm	1.45																																																																							
Continuous 0.5 lpm	3																																																																							
Continuous 1.0 lpm	1.45																																																																							
Continuous 1.5 lpm	1.15																																																																							
Continuous 2.0 lpm	0.45																																																																							

